

VERIFICATION OF CALIBRATION REPORT*of DataMaster cdm Breath Test Instrument**State of Alaska**Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*DEC 05 2010
[Signature]

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130116</u>	
Name <u>Kenneth Parker</u>		ID# <u>5578</u>	Date <u>11-19-2009</u>
A	Agency <u>McGrath Public Safety</u>		Phone # <u>524-3075</u>
Instrument Location <u>McGrath AST office</u>			
Alco S/N <u>77111</u>		Target Value <u>.078</u>	High Pressure <u>700</u>
Alco Test Values		<u>.072</u> ✓	<u>.072</u> ✓
		1 st Alco	2 nd Alco
Signature <u>[Signature]</u>			
		COB 1/9/10	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 14th day of January, 2010.

(Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130116

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

NOVEMBER 19, 2009

OPERATOR'S NAME:

PARKER/KENNETH/W

OPERATOR'S NUMBER: 5578

SUBJECT'S LAST NAME:

PARKER

SUBJECT'S FIRST NAME/MI :

KENNETH/W

O.L. #: 130116

DEPT/AGENCY: MC61

CASE/REPORT: T

TEST TYPE: T

ALCO TARGET VALUE: 078

ALCO S/N: 77111

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.68 in	.077	15:22
ALCO TARGET	.000	15:22
BLANK TEST	.000	15:22
INTERNAL STANDARD	VERIFIED	15:23
ALCO TV 29.68 in	.072	15:23
BLANK TEST	.000	15:24
SUBJECT SAMPLE	.000	15:25
BLANK TEST	.000	15:25
ALCO TV 29.68 in	.072	15:26
BLANK TEST	.000	15:27

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130116

NOVEMBER 19, 2009

TIME 15:17

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER:

49c

BREATH TUBE:

41c

BAROMETER:

29.68 in

FLOW DETECTOR:

OKAY

PUMP

HIGH SPEED:

OKAY

DETECTOR:

OKAY

FILTERS:

OKAY

QUARTZ STANDARD:

OKAY

CALIBRATION:

OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~